

CONDUCTIVITY BUFFER SOLUTIONS

Conductivity standard solutions certified by the Danish metrology DFM (referring ACCREDIA), credited for the measurement of conductivity, traceable to NIST. Expired date 18 months (5 microS) and 12 months (1.3 microS).

Glass bottle	300 ml
1,3 microS/cm @ 25°C	32389733
5 microS/cm @ 25°C	32389263



Conductivity standard solutions with certified DFM / ACCREDIA The product is valid for a period of 12-18 months.

Bottle with compartment AutoCal calibration built	500 ml
84 microS/cm @ 25°C	32389843
147 microS/cm @ 25°C	32389853
1413 microS/cm @ 25°C 1278 microS/cm @ 20°C KCl 0.01 M	32389863
12880 microS/cm @ 25°C, 11670 microS/cm @ 20°C KCl 0.1 M	32389883

Special conductivity standard solutions for the calibration of conductivity from plant and laboratory certified ISO 9001 AND INAB (referring ACCREDIA). Expired date 6-18 months, except the standard 10 microS, 3 months valid.

Bottle	500 ml
10 microS/cm @ 25°C	12006233
20 microS/cm @ 25°C	12006243
50 microS/cm @ 25°C	12006253
100 microS/cm @ 25°C	12006263
200 microS/cm @ 25°C	12006273
500 microS/cm @ 25°C	12006283
1000 microS/cm @ 25°C	12006293
5000 microS/cm @ 25°C	12006303
10.000 microS/cm @ 25°C	12006343
20.000 microS/cm @ 25°C	12006313
50.000 microS/cm @ 25°C	12006323
100.000 microS/cm @ 25°C	12006333
150.000 microS/cm @ 25°C	12006383
200.000 microS/cm @ 25°C	12006393



SATEMA

13856 VIGLIANO B.SE - Via Milano, 395

Tel. +39 015811102 - 015510156 Fax 0158853029

Mail: info@satema.it <http://www.satema.it>